
Incomplete ANA Screen Reports

Dear Regional Pathology Services clients,

We recently discovered that one of the components of the ANA Screen (Test Code ANASCR), dsDNA, is being performed but not reported when the ANA Screen is positive.

Until the test build issue is resolved, we will push the dsDNA results, triggering a corrected report.

We appreciate your understanding and patience as we work to correct the electronic results!

Contact Regional Pathology Services

Call: 800-334-0459 | email: pathclientservrps@unmc.edu